

Calibration and cross-laboratory implementation of scintillating screens for electron bunch charge determination

Kurz, T.; Couperus, J. P.; Krämer, J. M.; Ding, H.; Kuschel, S.; Köhler, A.; Zarini, O.;
Hollatz, D.; Schinkel, D.; D'Arcy, R.; Schwinkendorf, J. P.; Irman, A.; Schramm, U.;
Karsch, S.;

Originally published:

September 2018

Review of Scientific Instruments 89(2018), 093303

DOI: <https://doi.org/10.1063/1.5041755>

Perma-Link to Publication Repository of HZDR:

<https://www.hzdr.de/publications/Publ-27291>

Release of the secondary publication
on the basis of the German Copyright Law § 38 Section 4.

Calibration and cross-laboratory implementation of scintillating screens for electron bunch charge determination

Thomas Kurz,^{1,2,3, a)} Jurjen Pieter Couperus,^{1,3} Jakob Matthias Krämer,^{1,3} Hao Ding,^{2,4} Stephan Kuschel,^{5,6} Alexander Köhler,^{1,3} Omid Zarini,^{1,3} Dominik Hollatz,^{5,6} David Schinkel,^{5,6} Richard D'Arcy,⁷ Jan-Patrick Schwinkendorf,^{7,8} Jens Osterhoff,⁷ Arie Irman,¹ Ulrich Schramm,^{1,3} and Stefan Karsch^{2,4}

¹⁾ *Helmholtz-Zentrum Dresden-Rossendorf, Bautzner Landstraße 400, 01328 Dresden, Germany*

²⁾ *Ludwig-Maximilians-Universität München, Am Coulombwall 1, 85748 Garching, Germany*

³⁾ *Technische Universität Dresden, 01069 Dresden, Germany*

⁴⁾ *Max-Planck-Institut für Quantenoptik, Hans-Kopfermann-Straße 1, 85748 Garching, Germany*

⁵⁾ *Helmholtz-Institut Jena, Fröbelstieg 3, 07743 Jena, Germany*

⁶⁾ *Friedrich-Schiller-Universität Jena, Fürstengraben 1, 07743 Jena, Germany*

⁷⁾ *Deutsches Elektronen-Synchrotron DESY, Notkestraße 85, 22607 Hamburg, Germany*

⁸⁾ *Institut für Experimentalphysik, Universität Hamburg, 22761 Hamburg, Germany*

(Dated: 26.08.2018)

We revise the calibration of scintillating screens commonly used to detect relativistic electron beams with low average current, e.g. from laser-plasma accelerators (LPAs), based on new and expanded measurements that include higher charge density and different types of screens than previous work¹. Electron peak charge densities up to 10 nC/mm² were provided by focused picosecond-long electron beams delivered by the ELBE linear accelerator at the Helmholtz-Zentrum Dresden-Rossendorf. At low charge densities, a linear scintillation response was found, followed by the onset of saturation in the range of nC/mm². The absolute calibration factor (photons/sr/pC) in this linear regime was measured to be almost a factor of 2 lower than reported by Buck et al.¹ retrospectively implying a higher charge in the charge measurements performed with the former calibration. A good agreement was found with the results provided by Glinec et al.² Furthermore long-term irradiation tests with an integrated dose of approximately 50 nC/mm² indicate a significant decrease of the scintillation efficiency over time. Finally, in order to enable the transfer of the absolute calibration between laboratories, a new constant reference light source has been developed.

I. INTRODUCTION

Triggered by the proposal of Tajima and Dawson³ in 1979, the development of laser-plasma accelerators (LPA) has progressed tremendously. Continuous advancement in both the understanding and control over the electron injection and acceleration mechanisms^{4,5} as well as the development of state of the art laser-systems up to the petawatt-regime^{6,7}, enables the acceleration of quasi-monoenergetic⁸⁻¹⁰ electron bunches to energies in the GeV-range¹¹⁻¹³ and peak charges in the nC-range¹⁴ within only cm-long acceleration distance. Compared to electron beams from conventional RF-accelerators, LPA beams possess unique characteristics, such as ultra-short duration and high peak current¹⁵, but still suffer from shot-to-shot fluctuations. Moreover the high intensity laser-plasma interaction creates a harsh environment where most of the conventional diagnostics fail. Therefore an appropriate diagnostic method is demanding. In particular for the measurement of energy-resolved single shot electron bunch charge, a method capable of covering a broad parameter range is required. Typical

bunches contain a charge of a few pC up to nC within ps-to fs-duration with an energy distribution ranging from 10 MeV to the GeV-level. Typically this is measured by a permanent or current-driven dipole magnet of m-scale length, which maps the electron energy to position in the dispersive plane, in combination with a correspondingly large (few hundreds of cm²) scintillating screen imaged onto a CCD-camera for charge diagnostic⁷. The majority of scintillation screens used in this work consist of a 10 μ m to 100 μ m-thick layer of powdered rare earth phosphor (Gd₂O₂S:Tb), which converts a fraction of the deposited electron energy into visible light. This process is dominated by fluorescence and has a lifetime of approximately 1 ms¹⁶. The short life-time enables single-shot diagnostic at repetition rates of up to 1 kHz. In contrast, imaging plates¹⁷⁻²⁰, which deliver a good spatial resolution and high dynamic range, suffer from a long offline readout procedure. Scintillating screens are commercially available, often under the trade name LANEX, and marketed for X-ray detection. They are manufactured in various spatial resolutions and sensitivities and are supplied by several companies. Thus, the choice of screen depends on the specific experimental requirements. Aimed for x-ray detection, generally no electron-photon conversion efficiency is specified and, therefore, careful calibration is required in order to extract quantitative electron charge

^{a)}E-Mail adress: t.kurz@hzdr.de

density data from screen images.

For this reason, calibration studies of scintillators have been performed some years ago by Glinec et al.², Buck et al.¹ and Nakamura et al.²¹. Among those, the work of Buck et al.¹ performed by the consortium authoring this work provides an absolute and transportable charge calibration for various scintillating screens allowing different laboratories to use this calibration to compare the obtained electron bunch charge across different experiments. Here, we update and extend this work towards higher charge densities, which reflects recent improvements especially in LPAs. In this new parameter range, more emphasis has to be given to saturation and radiation damage effects. Additionally, we update the range of screen types according to current commercial availability. Thus, the setup (sec. II) has been revised compared to the earlier work¹, charge densities have been adjusted to better meet the conditions of modern high-charge LPAs and the range of the linear response (sec. III A) as well as the saturation (sec. III B) of the scintillating screens has been investigated. Previously unavailable information on the long-term stability and damage resistance under repeated high-flux irradiation is reported in section III C. In section IV, a new concept for the cross-calibration of the scintillating screens is described. This concept enables to implement the calibration results in other laboratories. This study is intended to replace the previous study by Buck et al.¹.

II. EXPERIMENTAL SETUP

The setup for the absolute charge calibration of the scintillation screens is illustrated in Fig. 1. The measurements were performed at the ELBE linear accelerator (LINAC) at the Helmholtz-Zentrum Dresden-Rossendorf. For this campaign, the sub-10 ps long electron bunches with a charge of 15 pC at an energy of 23.5 MeV were provided by the LINAC. In order to extend the calibration range to higher charges, the accelerator can be operated in a bunch train mode with multiple pulses in one train of tunable length running at a repetition rate of 13 MHz. The temporal spacing between the pulses corresponds to 77 ns. Since the lifetime of the excited state of the scintillator is ~ 1 ms¹⁶, the charge of 15 pC contained in each micro-bunch in a train of up to 1400 bunches adds up which in total form a macro-bunch with a charge of up to 21 nC. After focusing the electron beam by magnetic quadrupoles to a full width at half maximum (FWHM) area of 2 mm^2 to 3 mm^2 this corresponds to a peak charge density of up to 10 nC/mm^2 . Although measurements are performed at a fixed electron energy of 23.5 MeV, the calibration applies reliably over a broad energy range. Simulations show that the energy deposition of the electrons inside the photo-luminescent layer is almost independent of their kinetic energy above a threshold-value of 3 MeV^{2,18,22,23}. Additionally, Nakamura et al.²¹ demonstrated that the sensitivity of Lanex189

Fast to highly relativistic electrons have only a weak dependency of the electron energy. It decreases by 1% per 100 MeV increase of the electron energy. Thus the calibration results can be used to determine the charge of relativistic electron bunches produced by LPAs.

Directly before interacting with the screen, the charge of each electron micro-bunch is measured by an integrating current transformer (ICT-082-070-05:1-VAC, Bergoz Instrumentation, France). The ICT pulses were amplified by a factor of 56 (Pulse Amplifier Coaxial ZPUL-30P, Mini Circuits, USA) and recorded by a high resolution oscilloscope (2GHz RTE 1204, Rhode&Schwarz, Germany). The fluorescence emission from the scintillating screens exhibits an angular distribution following approximately the Lambertian law^{23,24}. The setup geometry was chosen such that the camera collects light in the direction normal to the screen which maximizes the light detected by the CCD-chip. The scintillating screens were placed within a vacuum vessel which was directly connected to the main accelerator. Furthermore a metallic turning mirror (PF20-03-P01, Thorlabs) was mounted off-axis to avoid generating background OTR that would otherwise contribute to the signal detected on the CCD-chip. These precautions improved on the setup used in the original work of Buck et al.^{1,19}. The screens were mounted on a rotating target wheel which was aligned $(22 \pm 1)^\circ$ relative to the electron beam. The emitted photons with a peak wavelength λ_{peak} of 546 nm are reflected by the mirror to a 12-bit CCD-camera (acA1300-30gm, Basler) equipped with a tele-objective (Ricoh FL-CC6Z1218A-VG, Ricoh, Japan) having a focal length of $(53 \pm 2) \text{ mm}$ and an f-number of 1.8. For the broad charge range treated in this work, a detection system with a dynamic range of three orders of magnitude is required. This is achieved by a filter wheel equipped with ND-filters ranging from ND0.5 to ND4.0 which were calibrated precisely (below 0.5% uncertainty) using a well-calibrated spectrophotometer (Cary 50 UV-VIS, Agilent Technologies). This approach guarantees a more precise measurement of the scintillator brightness (photons/sr) over the entire calibration curve, as compared to Buck et al.¹ where an adjustable iris was used. Additionally an optical fiber (M200L02S-A, Thorlabs) connected to a spectrometer (HR4000, Ocean Optics) was used to determine the fluorescence-spectrum of the scintillation screens. This information is required to determine the camera efficiency of the CCD and the effective attenuation of the ND-filters over the whole scintillation spectrum. In order to cross-reference the scintillation signal with a master light source, a tritium source was added to a free position on the filter wheel. The solid angle in our optical detection system was $(3.18 \pm 0.07) \times 10^{-3} \text{ sr}$ defined by an aperture diameter of $(22.96 \pm 0.05) \text{ mm}$ mounted at a distance of $(361 \pm 4) \text{ mm}$ from the target. For this small solid angle, the lateral signal variation due to the Lambertian emission characteristic can be ignored.




FIG. 1. Setup for absolute charge calibration of scintillation screens: The ICT measures the charge of the electron beam. Six different screens with an angle of 22° relative to the incoming electron beam were mounted on a rotating target wheel and optically imaged via a silver mirror onto a CCD-chip. In order to generate three orders of magnitude of dynamic range a set of ND-filters was placed in front of the camera. The effective collection angle is defined by an aperture in front of the telephoto lens.




FIG. 2. Image of an electron bunch recorded by the CCD-sensor. The rectangle marks the region of interest (ROI) which was used for the analysis. The two curves indicate the line-out of the electron bunch through its peak in horizontal and vertical direction. The area of the bunch at FWHM $\approx 2 \text{ mm}^2$.

III. RESULTS

A. Absolute charge calibration

The absolute calibration (total photons/sr/pC) of the scintillation screens serves as a universal reference for charge diagnostics with particular importance on the field of LPAs. Therefore the transmission efficiency of the optical detection system was measured to determine the brightness (photons/sr) of the scintillator. Together with

a precise knowledge (5% uncertainty) of the LINAC's bunch charge, the absolute scintillation efficiency can be determined.

A representative electron beam image is shown in Fig. 2. The brightness (photons) of the scintillating screens is measured as the integrated CCD-counts within the region of interest (ROI) and corrected for the background originating from the camera, i.e. thermal noise, the accelerator dark current and bremsstrahlung. Accordingly, the absolute response of the scintillating screen C_S , i.e. the total number of photons N_{ph} emitted by the scintillator into an area of one steradian per incident electron charge Q_e is given by

$$C_S = \frac{N_{\text{ph}}}{Q_e} = \frac{N_c \cos(\varphi)}{\beta \Omega Q_e}, \quad (1)$$

where N_c describes the total number of counts within the ROI of the background-corrected image. φ is the angle between the electron beam and the normal vector of the scintillator's surface. The cosine corrects the photon signal recorded by the CCD-camera for the increased interaction length due to the incidence angle of the electrons. Ω symbolizes the effective collection angle in units of steradian. Finally, β denotes the total efficiency of the entire detection system, i.e. the probability for a single photon that was emitted by the source, to travel through the optical system, reaching the CCD-chip and be converted to a count by the analog-to-digital converter. For the sake of completeness, β can be disassembled in its individual contributions. The reflection of the off-axis mirror at the specific wavelength is $(97 \pm 1)\%$, the window of the vacuum-chamber transmits $(91.3 \pm 0.5)\%$ of the incoming light and the objective transmits $(88 \pm 1)\%$ of the photons to the chip. The photon-to-count conversion efficiency of $(32.8 \pm 1.7)\%$ of the CCD-chip (ICX445, Sony) and its associated readout-electronics was determined separately using a green laser and a reference power detector (XLP12-3SH2-D0, Gentec International, Canada).

The response functions for the different screens as a function of the bunch charge are shown in Fig. 3. All curves show a linear behavior up to a threshold caused by saturation and degeneration effects (Sec. III B, III C). In order to determine the calibration value for the absolute response of the different scintillators, a linear fit has been applied to all data points that deviate less than -10% from linearity as indicated by the filled markers. As the upper limit of this linear region depends on the areal charge density, it obviously depends on the beam size (see section III B). The resulting calibration values are shown in the second column of Table I. It should be noted that only a single sample of each type of screen was investigated systematically. Potential batch-to-batch variations have not been investigated¹.

¹ We have irradiated the scintillation screens afterwards with a

TABLE I. Calibration values for different scintillation screens: Results for the absolute fluorescence efficiency (second column) and the saturation threshold (third column) as well as the resulting fit parameter (fourth column). The presented values reflect the absolute fluorescence efficiency and saturation behavior of fresh scintillating screens, not affected by irreversible damage effects. For comparison, the calibration values from Buck et al.¹ are listed for two screens also investigated in the earlier work.

Screen	Absolute fluorescence efficiency (10^9 ph/sr/pC)	Saturation threshold (10^3 pC/mm ²)	Birks' constant (10^{-5} mm ² /pC)
Kodak BioMAX MS	7.6 ± 1.3	1.6 ± 0.2	7.1 ± 0.8
Cawo OG BACK	5.8 ± 1.0	1.8 ± 0.2	6.0 ± 0.6
Cawo OG FRONT	3.7 ± 0.7	1.8 ± 0.2	6.1 ± 0.7
Konica Minolta OG 400	3.7 ± 0.7	1.9 ± 0.2	5.8 ± 0.6
Carestream Lanex Regular	3.1 ± 0.6	1.9 ± 0.2	5.9 ± 0.6
Kodak Lanex Fine	1.0 ± 0.2	3.5 ± 0.3	3.1 ± 0.4
Kodak BioMAX MS [Buck et al.]	14.8 ± 1.3		
Kodak Lanex Fine [Buck et al.]	1.75 ± 0.15		

The overall uncertainty of the absolute fluorescence efficiency of 16% (see Table I) originates from statistical and systematic uncertainties. The standard error of the mean (average over 50 shots) of the raw images together with the error of the linear fit add up to approximately 2%. The main contribution is due to systematic uncertainties. The error in the determination of the bunch charge Q_e was determined to be 5%, originating from the systematic error of the ICT measurement¹ and the uncertainty of the analysis algorithm. An error of 6% is accounted to the uncertainty of the transmission efficiency β , which includes optics in the beam-line (see Fig.1) as well as the photon-to-count efficiency of the CCD-camera. Furthermore, the cosine error caused by the deviation of the alignment angle φ and the solid angle Ω (see sec. II) add up to the remaining uncertainty of 3%. A detailed overview of all the relevant quantities for the determination of the absolute fluorescence efficiency can be found in Table II.

When comparing the absolute calibration results to Buck et al.¹, the sensitivity ratios between the screens agree quite well. However, a constant reduction of the overall sensitivity by roughly a factor of two is found in this work. The deviation might originate from two shortcomings in the setup of the previous work¹ which offer potential interfering background sources. First, in order to image the scintillating screen, a mirror was placed on the beam axis. As a source of OTR-light this mirror can add a significant amount of extra signal as partially collected by the camera²⁵. Secondly, the electron beam was coupled out from the vacuum chamber via a 100 μm thick beryllium window, potentially causing an extra charge-dependent X-ray background which the scintillating screen converts into light. The latter will add only a small amount of background signal, nevertheless in the

current calibration setup such an X-ray contribution was avoided.

The difference to Buck et al. prompted us to repeat the experiment in three independent campaigns in order to exclude sources of systematic errors. The setup for each campaign was identical apart from small (10%) variations in the collection angle, which have been taken into account in the analysis. The comparison of the absolute scintillating signal of each screen measured in the different calibration campaigns lead to similar ($\pm 5\%$) results increasing the confidentiality of the obtained calibration values. Additionally, it was found that the scintillation efficiency based on the experimental values published by Glinec et al.² shows good agreement to the current value. The data provided by Glinec leads to an absolute conversion efficiency for KODAK Lanex Fine of $(1.05 \pm 0.09) \times 10^9$ ph/sr/pC confirming the current calibration value of $(1.0 \pm 0.2) \times 10^9$ ph/sr/pC within the measurement uncertainty (see detailed derivation in the supplementary material). Consequently, the charge in LPA experiments has likely been underestimated in the past when the original calibration data of Buck et al.¹ was applied.

Moreover, we have investigated the responses of screens from an asymmetric cassette, e.g. Cawo OG F/B. For the original x-ray imaging application, scintillating screens are used to irradiate an analog film that is not sensitive to x-rays. In order to enhance the sensitivity, the scintillating screens are arranged in a sandwich-structure (screen-film-screen). Although most cassettes use two identical screens, some e.g. Cawo OG F/B are produced in a 'front' and a 'back' version, which are generally delivered in a single package. We found that the Cawo OG 'back'-side screen has an efficiency that is 50% higher than the efficiency of the 'front'-side screen.

broadband X-ray source to test the homogeneity of the investigated screen samples. We have not measured any significant variation along the screens. Additionally, a fresh screen-area was taken for each of the three campaigns to compare the absolute fluorescence efficiency resulting in very similar results.

B. Saturation effects

Beyond the linear region of the calibration curves, the photon-to-charge-conversion efficiency is reduced due to




FIG. 3. Absolute charge calibration of six different scintillation screens. The linearity hypothesis is valid up to a certain charge density threshold where the fluorescence signal is significantly lowered. We define the threshold at a drop in signal of more than 10% below the low-charge linear fit. Beyond this threshold, nonlinear saturation effects and beam induced damage effects start to play a role in the photon response. The solid line of each calibration curve indicates the linear region whereas the dotted line shows the non-linear part. The dashed line displays the calibration curve for Kodak Lanex Fine from Buck et al.¹. Additionally two reference data points for Kodak Lanex Fine are included. The red circle is determined by a calculation based on a Monte-Carlo-Simulation reported in Glinec et al.² as referenced in Buck et al. The red square was deduced from the full set of experimental results given by Glinec et al. (see supplementary material). The FWHM-area of the beam is $\sim 2 \text{ mm}^2$.

323 saturation in the active layer of the scintillator. Birks' ³⁴⁴
 324 law²⁶, describing the saturation of scintillators, is used ³⁴⁵
 325 to fit the response curve of the scintillator: ³⁴⁶

$$326 \quad \rho_{\text{scint}} = \frac{\rho_{\text{ICT}}}{1 + B\rho_{\text{ICT}}}, \quad (2) \quad 347 \quad 348 \quad 349$$

327 where the fit parameter B is Birks' constant. Here, ρ_{ICT} ³⁵⁰
 328 is the applied peak charge density which is determined ³⁵¹
 329 by the electron bunch charge measured by the ICT and ³⁵²
 330 the beam profile of the scintillator in the linear region. ³⁵³
 331 Assuming a constant beam shape, we calculated ρ_{ICT} in ³⁵⁴
 332 the saturated regime using the charge information given ³⁵⁵
 333 by the ICT. ρ_{scint} is the peak charge density measured ³⁵⁶
 334 on the scintillator. The saturation threshold value ρ_{sat} is ³⁵⁷
 335 defined as the peak charge density, at which the scintil-
 336 lation signal has dropped to 90% compared to the linear ³⁵⁸
 337 behavior. This arbitrary measure is chosen such that the ³⁵⁹
 338 saturation effect can be clearly distinguished from rela- ³⁶⁰
 339 tive measurement uncertainties in the linear case. Fig. ³⁶¹
 340 4 shows a saturation curve of Kodak BioMAX MS with ³⁶²
 341 increasing electron peak charge density. The dashed line ³⁶³
 342 shows the linear correlation of ρ_{scint} and ρ_{ICT} , while the ³⁶⁴

solid curve indicates the fit along the measured data. The resulting threshold values and the fit parameter B for the different screens are shown in Table I. It should be noted that the experimental implementation of the setup potentially underestimates this effect. For the highest applied charges, the duration of the pulse train is in the order of 100 μs and becomes comparable to the lifetime of the excited state. Thus, electron bunches in the tail of the bunch-train have an enhanced probability to re-excite atoms that have already relaxed back and thus add less to saturation. This effect has been included in Fig. 4 as an increased uncertainty towards lower scintillation peak charge density and is only relevant for the last two data points.

At high peak charge densities, all investigated scintillating screens start to saturate. A weakly focused electron beam was used to increase the peak charge density by more than two orders of magnitude compared to previous saturation studies¹. In contrast to Ref. 1, we observe saturation of the scintillator starting when applying peak charge density on the order of nC/mm^2 .




FIG. 4. Response function of Kodak Biomax MS showing saturation: The peak charge density measured by the screen vs. the peak charge density calculated from the beam profile of the scintillator and the charge information given by the ICT. The bunch profile shows a significant saturation towards higher charges. The measured data is fitted with Birks' law of saturation (solid line, see eq. 2). The black dotted line indicates $\rho_{\text{scint}} = \rho_{\text{ICT}}$.

C. Degradation effects

Besides reversible saturation, additional and previously unreported beam-induced degradation occurs. In order to separate saturation from degradation damage, reference measurements were performed with a low charge of 60 pC after each increment of the bunch charge to get a reasonable estimate for the degeneration caused by the increasing charge densities applied during the calibration measurements. Typically, it took several minutes to switch to the reference measurement. The beam-induced fluorescence reduction is defined as the reduction of the fluorescence signal after a recovery time of several minutes. Due to a limited access time to the accelerator, the recovery behavior of the screens could not be studied at longer timescales. Nevertheless this effect becomes relevant when frequently irradiating the scintillating screens with relativistic electrons i.e. LPA experiments performed with Hz-repetition rate for several hours (see Fig. 5). After applying the cumulative charge of the entire calibration curve, the beam-induced damage by the measurement itself is consistently circa 15% for all screens except Kodak Lanex Fine, which did not show any decay. Again, this effect becomes relevant only at high electron doses, i.e. the last 2–3 data points in Fig. 4. The values in Table I (third & fourth column) and Fig. 4 are corrected for this damage by adding the reduction caused by the measurement to the measured signal.

For a reliable bunch charge diagnostic the long-term stability and possible degradation of the scintillating

screens has to be quantified. Already the calibration curves in Fig. 3 include some non-reversible degradation with accumulated dose confirmed by intermediate low-charge reference shots. In order to further investigate this apparent beam-induced degradation, we conducted two dedicated long-term-irradiation tests. These tests were independent from the calibration measurements and served as a different method than the above mentioned one to study irreversible damage effects in a much more controlled way. For each one we used a fresh sample of one specific type of scintillating screen i.e. Konica Minolta. These tests were performed with a constant charge density per shot far below the saturation threshold to exclude reversible saturation during the long-term-tests. The electron beam parameters were chosen to represent LPA-conditions as close as possible. Every second, the screen was irradiated by an electron bunch with a charge of 100 pC over a duration of 90 min. The FWHM-bunch area was kept at 2 mm² yielding electron densities at the target on the order of 9 pC/mm². In the following, results from two runs with almost equal beam parameters are described. Fig. 5 shows the fluorescence signal as a function of the applied cumulated electron charge density over time. The data was fitted with an exponential decay function with an offset at around 0.9 to determine the loss of scintillation efficiency. A significant drop of 9% in the emitted scintillation efficiency over 1.5 h was observed. In such a case, the charge will be underestimated accordingly. The influence of the repetition rate and the charge density on the long-term stability i.e. the decrease of the scintillation signal per total electron dose, was not investigated and should be studied in future work.

Fig. 6 displays the temporal evolution of the efficiency during a second long-term test, performed under similar conditions as in the first run. First, the scintillator also shows a decay as observed in Fig. 5. The beam profile for a representative shot onto the scintillator at an integrated dose of 50 nC/mm² is illustrated in Fig. 6. At a cumulative charge density of around 52 nC/mm², the response function shows a sharp increased peak at which the screen lights up brightly with a hole in its center. Afterwards the screen is permanently damaged. Due to the long irradiation time needed to cause this effect, we could not investigate how reproducible this sudden damage is, but it is worth recording its occurrence.

IV. IMPLEMENTATION OF THE LANEX CALIBRATION AT EXTERNAL LABORATORIES

Here we explain how potential users can implement the presented calibration results to measure the charge of relativistic electron bunches.

The values in Table I give the absolute light output into a small solid angle around the forward direction as a function of incident charge. In principle, these values could be used to directly obtain charge figures from absolutely measured photon numbers. The problem with this




FIG. 5. Long-term performance test with Konica Minolta. The screen was irradiated for 1.5 h with 1 Hz repetition rate, 100 pC charge and a spot size of 2 mm² at FWHM. The data was fitted with an exponential decay function that has an offset value at approximately 0.9. The decay of the photon signal during this experiment was 9%.




FIG. 6. Damage of Konica Minolta during long-term test: The data was taken at a different run with equal parameters as presented in Fig. 5. After applying a cumulative dose of 52 nC/mm², the screen shows a bright peak and is permanently damaged afterwards.

brute force approach is that it requires a precise knowledge of the absolute transmission of the imaging system and the camera response. The total photons emitted by the scintillator per steradian N_{Total} can be determined as :

$$N_{\text{Total}} = \frac{N_{\text{Counts}}}{\Omega \eta}, \quad (3)$$

where N_{Counts} is the absolute amount of counts detected by the CCD-camera, Ω denotes the collection angle. The efficiency of the specific optical detection system η is given by:

$$\eta = \eta_{\text{qe}} \cdot t_{\text{obj}} \cdot t_{\text{window}} \cdot r_{\text{mirror}}. \quad (4)$$

Hereby, η_{qe} is the camera efficiency, t_{obj} and t_{window} symbolize the transmission value through the objective and the vacuum window respectively and r_{mirror} is the reflectivity of the turning mirror. Whereas N_{Counts} and Ω are easy to determine, η_{qe} is rather complex. Each element (e.g. ND-filter, band-pass filter, mirror, vacuum window, camera objective etc.) of the optical detection system has to be calibrated, i.e. the spectral transmission has to be multiplied with the spectrum of the scintillator, which is similar for all screen types (an example spectrum of Kodak BioMax MS is given in the supplementary material). Special care is needed for the calibration of the camera, as the efficiency of the CCD-camera (photon-to-count) is not equal to the quantum efficiency (probability for an incident photon to create an electron-hole pair) of the CCD-chip but also depends on the transformation ratio from the analog CCD-signal to the amount of digital counts. This approach has the clear disadvantage that every change in the setup and the optical detection system directly affects the charge determination. Therefore a precise knowledge of each parameter in equation 3 is required. Additionally, the error in determining the signal of the screen (photons/sr) adds to the calibration error reported in Table I.

These problems can be overcome by calibrating the optical detection system with a constant-brightness reference light source (called master light source, MLS in the following). This light source is directly compared with the signal of the scintillating screen through the same optical system as during the LINAC-based screen calibration procedure. By applying the MLS specific charge density Q/A_{MLS} and taking into account the measured ratio between the screen B_{screen} and the MLS brightness B_{MLS} an to be measured charge density Q/A can be expressed as:

$$Q/A = \frac{B_{\text{screen}}}{B_{\text{MLS}}} Q/A_{\text{MLS}} \quad (5)$$

regardless of the imaging system. Q/A_{MLS} depends on the master light source and was obtained during the calibration campaign. However, due to its rather large geometrical dimensions and the vacuum incompatibility the source isn't implemented directly. Instead, it is used to cross-calibrate any amount of daughter light sources (DLS) in a standalone setup (see right image in Fig. 7) for distributing the calibration onto different experiments, where such a DLS is placed at the position of or

TABLE II. List of typical input parameters for the absolute calibration

Quantity	Symbol	Method of determination
Counts	N_{counts}	Sum over counts in the image
Solid angle	Ω	$\frac{A}{r^2}$; $\frac{\text{aperture}}{\text{distance to screen}^2}$
Efficiency	η	Transmission: green laser + power head Quantum efficiency: Sum of counts vs. Sum of photons by power head
$\left\{ \begin{array}{l} \text{Camera efficiency} \\ \text{Transmission objective} \\ \text{Transmission window} \\ \text{Reflectance mirror} \\ + (\text{more items}) \end{array} \right.$	$\left\{ \begin{array}{l} \eta_{\text{qe}} \\ t_{\text{obj}} \\ t_{\text{window}} \\ r_{\text{mirror}} \\ + (\text{more items}) \end{array} \right.$	

506 next to the scintillating screen. A charge density Q/A ₅₃₆
507 can then be determined via:

$$508 \quad Q/A = \frac{B_{\text{screen}}}{B_{\text{MLS}}} \frac{B_{\text{MLS}}}{B_{\text{DLS}}} Q/A_{\text{MLS}}, \quad (6)$$

509 where $B_{\text{MLS}}/B_{\text{DLS}}$ is the cross-calibration factor be-
510 tween the MLS and the DLS.

511 In the past, calibrated gaseous tritium light sources
512 (GTLS) were used as MLSs/DLSs due to their small size
513 and supposedly well-predictable light output. However,
514 their main drawback beyond a certain radiation hazard,
515 is an aging behaviour that is not synchronous with the
516 well-known Tritium decay rate, but also depends on an
517 unknown degradation of their luminous phosphor. This
518 makes it impossible to confidently predict and correct
519 for the degradation of the light output over time. There-
520 fore, for the use as MLS the Tritium capsules would have
521 to be calibrated in a LINAC beamtime typically every
522 year. The obvious solution for this problem is to design
523 a truly constant MLS that is by far more accessible than
524 a LINAC beam-time. Such a source can be found at
525 the Helmholtz-Zentrum Dresden-Rossendorf² or at the
526 Ludwig-Maximilians-Universität München.

527 Finally, this method eliminates most error sources and
528 enables the transport of calibration results to other lab-
529 oratories. The new LED-based MLS is an improvement
530 over the previously reported MLS¹, because the amount
531 of emitted light is stable over time. In certain scenarios,
532 GTLS, implemented as DLS, are still preferred due to
533 their small size and vacuum compatibility. The overall
534 error of the charge density determination using a cali-
535 brated DLS will be in the order of 10%.

537 V. CONCLUSION

538 We have presented an absolute charge calibration mea-
539 surements, saturation effect study and long-term stabil-
540 ity tests for various scintillating screens that are com-
541 monly used as the diagnostic for short-pulsed electron
542 bunches with low average current and spreaded over a
543 rather large area as in the case of LPAs. The measure-
544 ments were performed with sub 10 ps, 23.5 MeV elec-
545 tron beams from the ELBE linear accelerator at the
546 Helmholtz-Zentrum Dresden-Rossendorf. The absolute
547 scintillation efficiency of the different scintillation screens
548 was measured to vary over almost one order of mag-
549 nitude. Kodak BioMAX MS shows the brightest pho-
550 ton response with an absolute fluorescence efficiency of
551 $(7.6 \pm 1.3) \times 10^9$ photons/sr/pC but also is the screen
552 which saturates first. For comparison, we have calcu-
553 lated the absolute response of Kodak Lanex Fine based
554 on the experimental data provided by Glinec et al.². The
555 resulting value of $(1.05 \pm 0.09) \times 10^9$ ph/sr/pC agrees to
556 our value of $(1.0 \pm 0.2) \times 10^9$ ph/sr/pC within the mea-
557 surement uncertainties.

558 A saturation effect was visible when applying peak
559 charge densities in the order of nC/mm². This is about
560 three orders of magnitude higher than charge densities
561 reached in current LPA experiments and can therefore
562 generally be neglected when analyzing the spectrometer
563 images.

564 Finally the long-term stability for a selected type of
565 screen, i.e. Konica Minolta, was tested. We show that
566 a typical electron dose, i.e. 9 pC/mm², applied for 1.5 h
567 at 1 Hz leads to a significant decrease of the fluorescence
568 efficiency. Additionally, we found that heat damage of
569 LANEX screens becomes an issue after prolonged con-
570 tinuous use. Thus a careful heat dissipation concept has
571 to be established before implementing those screens in
572 accelerators with continuous operation mode.

573 The absolute calibration factors given in this work have
574 the disadvantage that they require a precise calibration
575 of the screen imaging system's spectrally and spatially
576 dependent throughput. In order to facilitate the calibra-
577 tion, we offer a relative screen calibration based on abso-
578 lutely calibrated reference daughter light sources. Since
579 the GTLS' luminosity is not stable over time, a new con-
580 cept for the cross-calibration of the scintillating screens

² A cross-calibration measurement for daughter light sources can
be performed by the authors including this work into the refer-
ence list of relevant publications making use of the calibration.
The MLS is based on a constant current source driving a green
LED with a specified lifetime of more than 10 000 hours. Due
to its infrequent use, this translates to a practical usability over
several decades.




FIG. 7. Schematic overview for the implementation of the calibration in the electron diagnostic in an external LPA. Step 1 is the calibration of the scintillation screens and the cross calibration to a master LED in a RF-accelerator facility as described in this article (see also Fig. 1). Second step is the cross calibration of daughter lights source e.g. a GTLS or a small and vacuum compatible LED. Subsequently, these calibrated daughter sources can be implemented in any laboratory.

580 was developed. The MLS/DLS approach simplifies the 593
 581 application and transportation of this calibration study 594
 582 significantly. Additionally, the uncertainty of the charge 595
 583 determination across laboratories can be minimized be- 596
 584 low 10% if a carefully calibrated DLS is implemented. 597
 598

585 SUPPLEMENTARY MATERIAL

586 See supplementary material for the calculation of the 600
 587 reference points in Fig. 3 and the spectrum of Kodak 601
 588 BioMAX MS. 602
 603
 604
 605

589 ACKNOWLEDGMENTS

590 This work was supported by EC Horizon 2020 609
 591 LASERLAB-EUROPE/LEPP (Contract No. 654148), 610
 592 DFG through Munich Center of Advanced Photon- 612

ics (Exc 158) and SFB Transregio TR18 funding
 schemes, EU through H2020 project Euratom (Contract
 No.633053), MPG through International Max-Planck Re-
 search School of Advanced Photon Science (IMPRS-
 APS), the Helmholtz Association ARD research topic
 and the Virtual Institute VH-VI-503.

599 ¹A. Buck, K. Zeil, A. Popp, K. Schmid, A. Jochmann, S. D.
 600 Kraft, B. Hidding, T. Kudyakov, C. M. S. Sears, L. Veisz,
 S. Karsch, J. Pawelke, R. Sauerbrey, T. Cowan, F. Krausz, and
 U. Schramm, *Rev. Sci. Instrum.* **81**, 033301 (2010).

602 ²Y. Glinec, J. Faure, A. Guemnie-Tafo, V. M. Monard, J. P. Lar-
 603 bre, V. De Waele, J. L. Marignier, M. Mostafavi, V. Malka, and
 H. Monard, *Rev. Sci. Instrum.* **77**, 103301 (2006).

604 ³T. Tajima and J. M. Dawson, *Phys. Rev. Lett.* **43**, 267 (1979).

605 ⁴E. Esarey, C. B. Schroeder, and W. P. Leemans, *Rev. Mod.*
 606 *Phys.* **81**, 1229 (2009).

607 ⁵M. C. Downer, R. Zgadzaj, A. Debus, U. Schramm, and M. C.
 608 Kaluza, *Rev. Mod. Phys.* **90**, 035002 (2018).

609 ⁶E. W. Gaul, M. Martinez, J. Blakeney, A. Jochmann,
 M. Ringuette, D. Hammond, T. Borger, R. Escamilla, S. Dou-

- 613 glas, W. Henderson, G. Dyer, A. Erlandson, R. Cross, J. Caird,⁶⁴⁹
614 C. Ebberts, and T. Ditmire, *Appl. Opt.* **49**, 1676 (2010).⁶⁵⁰
- 615 ⁷U. Schramm, M. Bussmann, A. Irman, M. Siebold, K. Zeil, D. Al-⁶⁵¹
616 bach, C. Bernert, S. Bock, F. Brack, J. Branco, J. Couperus,⁶⁵²
617 T. Cowan, A. Debus, C. Eisenmann, M. Garten, R. Gebhardt,⁶⁵³
618 S. Grams, U. Helbig, A. Huebl, T. Kluge, A. Köhler, J. Krämer,⁶⁵⁴
619 S. Kraft, F. Kroll, M. Kuntzsch, U. Lehnert, M. Loeser, J. Met-⁶⁵⁵
620 zkes, P. Michel, L. Obst, R. Pausch, M. Rehwald, R. Sauerbrey,⁶⁵⁶
621 H. Schlenvoigt, K. Steiniger, and O. Zarini, *J. J. Phys. Conf.*⁶⁵⁷
622 *Ser.* **874**, 012028 (2017).⁶⁵⁸
- 623 ⁸C. G. R. Geddes, C. S. Toth, J. Van Tilborg, E. Esarey, C. B.⁶⁵⁹
624 Schroeder, D. Bruhwiler, C. Nieter, J. Cary, and W. P. Leemans,⁶⁶⁰
625 *Nature* **431**, 538 (2004).⁶⁶¹
- 626 ⁹J. Faure, Y. Glinec, A. Pukhov, S. Kiselev, S. Gordienko,⁶⁶²
627 E. Lefebvre, J.-P. Rousseau, F. Burgy, and V. Malka, *Nature*⁶⁶³
628 **431**, 541 (2004).⁶⁶⁴
- 629 ¹⁰S. P. D. Mangles, C. D. Murphy, Z. Najmudin, a. G. R. Thomas,⁶⁶⁵
630 J. L. Collier, a. E. Dangor, E. J. Divall, P. S. Foster, J. G. Gal-⁶⁶⁶
631 lacher, C. J. Hooker, D. a. Jaroszynski, a. J. Langley, W. B.⁶⁶⁷
632 Mori, P. a. Norreys, F. S. Tsung, R. Viskup, B. R. Walton, and⁶⁶⁸
633 K. Krushelnick, *Nature* **431**, 535 (2004).⁶⁶⁹
- 634 ¹¹C. B. Schroeder, C. Tóth, B. Nagler, a. J. Gonsalves, K. Naka-⁶⁷⁰
635 mura, C. G. R. Geddes, E. Esarey, S. M. Hookert, and W. P.⁶⁷¹
636 Leemans, *Conf. Proc. - Lasers Electro-Optics Soc. Annu. Meet.*⁶⁷²
637 **2**, 538 (2007).⁶⁷³
- 638 ¹²X. Wang, R. Zgadzaj, N. Fazel, Z. Li, S. A. Yi, X. Zhang,⁶⁷⁴
639 W. Henderson, Y.-Y. Chang, R. Korzekwa, H.-E. Tsai, C.-H.⁶⁷⁵
640 Pai, H. Quevedo, G. Dyer, E. Gaul, M. Martinez, a. C. Bern-⁶⁷⁶
641 stein, T. Borger, M. Spinks, M. Donovan, V. Khudik, G. Shvets,⁶⁷⁷
642 T. Ditmire, and M. C. Downer, *Nat. Commun.* **4**, 1988 (2013).⁶⁷⁸
- 643 ¹³W. P. Leemans, a. J. Gonsalves, H. S. Mao, K. Nakamura,⁶⁷⁹
644 C. Benedetti, C. B. Schroeder, C. Tóth, J. Daniels, D. E. Mittel-⁶⁸⁰
645 berger, S. S. Bulanov, J. L. Vay, C. G. R. Geddes, and E. Esarey,⁶⁸¹
646 *Phys. Rev. Lett.* **113**, 1 (2014).⁶⁸²
- 647 ¹⁴J. P. Couperus, R. Pausch, A. Köhler, O. Zarini, J. M. Krämer,⁶⁸³
648 M. Garten, A. Huebl, R. Gebhardt, U. Helbig, S. Bock, K. Zeil,
A. Debus, M. Bussmann, U. Schramm, and A. Irman, *Nat. Com-
mun.* **8**, 487 (2017).
- ¹⁵O. Lundh, J. Lim, C. Rechatin, L. Ammoura, A. Ben-Ismaïl,
X. Davoine, G. Gallot, J.-P. Goddet, E. Lefebvre, V. Malka, and
J. Faure, *Nat. Phys.* **7**, 219 (2011).
- ¹⁶R. Morlotti, M. Nikl, M. Piazza, and C. Boragno, *J. Lumin.*
72-74, 772 (1997).
- ¹⁷K. A. Tanaka, T. Yabuuchi, T. Sato, R. Kodama, Y. Kitagawa,
T. Takahashi, T. Ikeda, Y. Honda, and S. Okuda, *Rev. Sci.
Instrum.* **76**, 013507 (2005).
- ¹⁸S. Masuda, E. Miura, K. Koyama, and S. Kato, *Rev. Sci. In-
strum.* **79**, 083301 (2008).
- ¹⁹K. Zeil, S. D. Kraft, A. Jochmann, F. Kroll, W. Jahr,
U. Schramm, L. Karsch, J. Pawelke, B. Hidding, and G. Pretzler,
Rev. Sci. Instrum. **81**, 013307 (2010).
- ²⁰T. Bonnet, M. Comet, D. Denis-Petit, F. Gobet, F. Hannachi,
M. Tarisien, M. Versteegen, and M. M. Aleonard, *Rev. Sci. In-
strum.* **84**, 013508 (2013).
- ²¹K. Nakamura, A. J. Gonsalves, C. Lin, A. Smith, D. Rodgers,
R. Donahue, W. Byrne, and W. P. Leemans, *Phys. Rev. Accel.
Beams* **14**, 062801 (2011).
- ²²B. Hidding, G. Pretzler, M. Clever, F. Brandl, F. Zamponi,
A. Lübcke, T. Kämpfer, I. Uschmann, E. Förster, U. Schramm,
R. Sauerbrey, E. Kroupp, L. Veisz, K. Schmid, S. Benavides, and
S. Karsch, *Rev. Sci. Instrum.* **78**, 083301 (2007).
- ²³Y. C. Wu, B. Zhu, K. G. Dong, Y. H. Yan, and Y. Q. Gu, *Rev.
Sci. Instrum.* **83**, 026101 (2012).
- ²⁴G. E. Giakoumakis and D. M. Miliotis, *Phys. Med. Biol.* **30**, 21
(1985).
- ²⁵B. Yang, “A design report for the optical transition radiation
imager for the LCLS undulator,” *Tech. Rep.* (SLAC, Stanford,
2005).
- ²⁶J. B. Birks, D. Fry, L. Costrell, and K. Kandiah, *The Theory
and Practice of Scintillation Counting* (Pergamon Press, Oxford,
1964).